



## AutoUA® Automated Urinalysis - Sample Patient Report

Sample ID:	DTC0001			Sex:	M
Report Date/Time:	6/25/2025 17:09			Age(Years):	69
Sample Type:	Urine				
Instrument:	AUA-680				

Test Information					
Test	Quantitative	Normal Range (Low/High)		Normalized	Qualitative
Leukocyte (LEU/L)	3.8	0.0	3.0	9.6	Trace
Urobilinogen (mg/dL)	0.7	0.1	1.9	1.8	Normal
pH (Units)	6.0	4.8	8.0	N/A	N/A
Specific Gravity (Units)	1.010	1.002	1.035	N/A	N/A
Bilirubin (mg/dL)	0.0	0.0	1.4	0.0	Negative
Nitrite (mg/dL)	0.00	0.00	0.10	0.00	Negative
Total Protein (micro) (mg/dL)	4.2	0.0	14.0	10.6	Negative
Hemoglobin (Blood) (µg/dL)	0.0	0.0	500.0	0.0	Negative
Ketones (mg/dL)	5.6	0.0	9.9	14.2	Negative
Glucose (mg/dL)	3.2	0.0	15.0	8.1	Negative
Creatinine (mg/dL)	49.9	20.0	500.0	N/A	N/A
Albumin (micro) (mg/dL)	0.0	0.0	1.9	0.0	Normoalbuminuria

Factors*	Normal Range (Low/High)		Quantitative	Normalized
Non-Albumin Proteins (NAP) (mg/dL)	0	2.0	4.2	10.6
(µ) Total Urine Protein:Creatinine Ratio (µPCR) (mg/g)	0	150	84.2	213.2
(µ) Albumin:Creatinine Ratio (µACR) (mg/g)	0	30	0.0	0.0
Non-Albumin Protein (NAP) Creatinine Ratio (mg/g)	0	120	84.2	213.2
Estimated 24 Hour Urinary Protein Excretion (UPEE) (g/day)	0	0.2	0.1	0.2

Note: NAP = Non Albumin Protein. Testing for only urine (µ) Albumin can miss up to 40 % of people with gross proteinuria caused by NAP!

\* The normal range for the "Factors" have been obtained from publications, peer-reviewed literature, and other sources.

All ranges should be verified as appropriate for your laboratory. Clinicians should review all results.

Clinicians should review all results.